



Adaptive Sampling With Trigger Conditions

5/22/14 Brent Roman brent@mbari.org



Traditional ESP Missions

- A sequence of “phases”, each with a prescribed start time
 - Actions predetermined by puck load
- ESP sleeps between phases. While “asleep”:
 - Still monitors contextual sensors
 - Still maintains radio context with shore
- All phases began at times prescribed in the mission script
 - Start times specified may be absolute or relative
 - Relative times specify the “sleep time” between phases
- No adaptive sampling was possible without hand coding it



Trigger Condition Overview

- Each start time is augmented by a list of trigger conditions
 - A phase starts when any of its trigger conditions is true
 - The start time can be thought of as the one required trigger condition
 - It determines the latest possible starting time for the phase
 - Triggers start phases before their scheduled times
 - Triggers cannot delay phases beyond their “start times”
 - Triggers **cannot** change the sequence of actions performed
 - » *Processing sequence is determined by puck load.*
- Each trigger condition is reevaluated whenever contextual sensors read
 - Sensible, as trigger conditions almost always evaluate sensor data
 - This is a convention
(but, not difficult to circumvent if necessary)
- Each trigger condition runs in its own Ruby thread
 - Failure (e.g. exceptions raised) in any trigger will not affect the others
 - You can even patch the code and restart failed trigger conditions
 - Or, kill the trigger thread to ensure it does not trigger the phase

Basic Trigger Conditions

- Basic Trigger Conditions contain arbitrary true/false expressions
 - A threshold value is associated with each
 - `CTD.temp < threshold`
 - `ISUS.no3 > threshold`
 - `CTD.depth > threshold[0]` and `CTD.fluor > threshold[1]`
 - Thresholds need not be scalar values
 - Trigger expressions are reevaluated just after each time contextual sensors are read while the mission is awaiting conditions
- May be assigned names like `Cold`, `Hot`, `Fresh`, `Salty`
- Threshold values can be modified at any time
 - Via the script itself or the interactively via `espclient`
 - All modifications to thresholds are logged
- Very flexible, but also painfully verbose for complex triggers

Composite Trigger Conditions

- Two types
 - Trigger “all” means when all subordinate conditions are true
 - **Trigger all: [Cold, DCM, HighNitrate]**
 - Equivalent to: **Cold[] and DCM[] and HighNitrate[]**
 - **Trigger all: []**
 - is always true
 - Trigger “any” means when any subordinate condition is true
 - **Trigger any: [Cold, DCM, HighNitrate]**
 - Equivalent to: **Cold[] or DCM[] or HighNitrate[]**
 - **Trigger any: []**
 - is always false
- All subordinate conditions run in the same thread as the parent

Trigger.now!

- Not really a trigger condition, rather an action!
 - Causes the current mission phase to start immediately
 - Raises an exception if mission is not waiting
 - Exception is raised in caller's thread
 - The mission's processing is unaffected
- There need not be any trigger conditions associated with the waiting phase for **Trigger.now!** to work.
 - The phase may be just awaiting its start time

Trigger.replace or Trigger.restart

- Replace current phase's start time and/or trigger conditions
 - Affects only for the phase currently waiting to start
 - Raises an exception if mission is not waiting
- All arguments are optional
- First argument is the replacement phase start time
 - Specify nil to leave start time unchanged
- Other arguments are replacement trigger conditions
 - Omit other args to leave existing triggers in place
- **Trigger.replace “+1.5 days”, Cold, Deep**
 - Mission will continue waiting up to 36 more hours for the **Cold or Deep** condition to be satisfied



Trigger Range Conditions

- True if each listed measurement is within one of the associated ranges of interest
 - Represented as a Ruby hash mapping keys to values. In this case:
 - Keys are measurements, like:
 - `CTD%:temperature`
 - `ISUS%:no3`
 - Not `CTD.temperature` #NO!!!
because that would check temperature once, when the condition was defined.
 - Values are ranges, like:
 - `-3.3 .. 2.1`
 - Trigger range:
`{CTD%:temperature => [-3.3..2.1, 5..7.21],`
`CTD%:salinity => [33..33.4, 23..28.3, 35..35.5],`
`ISUS%:no3 => 32.03..12.3}` #may omit [] for an array of one element
 - If first > last, as in `32.03..12.3` above, range check is logically negated
Equivalent to: `(no3 > 32.03 or no3 < 12.3)`

Trigger Box Conditions

- True if each listed measurement is within *the same* associated box of interest
 - Represented as the same Ruby hash mapping used for Trigger Ranges
 - Trigger box:
`{CTD%:temperature => [-3.3..2.1, 5..7.21],
CTD%:salinity => [33..33.4, 35..35.5]}`
 - Read the boxes off the columns of the resulting matrix.
 - If temperature is in one column and salinity is in the other, the trigger condition is *false*
- Columns geometrically define a set of boxes in the space of sensor measurements

Trigger Box Corner Cases

- If measurements do not specify the same number of ranges:
 - Those that are missing ranges will be ignored

Trigger box:

```
{CTD%:temperature => [-3.3..2.1, 5..7.21],  
CTD%:salinity      => [ 33..33.4 ]}
```

- If the temperature is between **5..7.21**, the trigger condition is true, regardless of salinity
- If a measurement specifies a single range (not an Array)
 - That range will be applied to all others
 - As though it had been repeated in an Array

Trigger box:

```
{CTD%:temperature => [-3.3..2.1, 5..7.21],  
CTD%:salinity      => 33..33.4}
```

- The salinity must always be in **33..33.4**, regardless of temperature

Trigger Holdoffs

- Trigger holdoffs are a simple way to avoid false triggers
 - A form of glitch filtering
 - ESP logs show countdown when awaiting holdoffs
- All triggers have an associated holdoff in samples
 - condition must be true for at least holdoff+1 samples
 - nil is the default holdoff value
 - holdoff=nil, equivalent of holdoff=0
 - But holdoff nil is not displayed, whereas 0 is
 - holdoff of false disables that particular trigger condition
 - holdoff of true forces trigger on its next evaluation

Trigger Thresholds

- Each trigger optionally has an associated threshold value
 - Usually used to parametrize conditional expressions
 - But you may choose to compare to constants instead
 - Need not be scalar, only the expression interprets it
 - Not usually applicable to box or range conditions
 - Such thresholds would be vectors of ranges if used
- If your conditional expressions reference a threshold:
 - You must set it before the trigger is used
 - `Cold.threshold = 4.3` #it's that easy!
 - The default threshold value is nil
 - `CTD.fluor > nil` #will generate an exception!

Trigger enable and disable

- Enable trigger monitoring with:
 - `Trigger.enable`
- Disable trigger monitoring with:
 - `Trigger.disable`
- Trigger monitoring is initially disabled
 - Use `Trigger.enable` as soon as contextual data starts making sense and all relevant thresholds are defined
- Triggers may be enabled/disabled at any time
 - Even while awaiting them
- Triggers are initially enabled during simulation!

Automatic Trigger Rearm

- Trigger monitoring may be disabled whenever a trigger condition causes a phase to start
 - If triggers remain enabled, rearm is said to be true
 - If triggers disable once one has fired, rearm is said to be false
- Set the rearm flag with:
 - `Trigger.rearm = true`
- Clear the rearm flag with:
 - `Trigger.rearm = false`
- Real missions start with `rearm=false`
 - You may change the `Trigger.rearm` flag at any time
 - You may want to combine it with `Trigger.enable` or `Trigger.disable`
- Simulation missions start with `Trigger.rearm=true`